

119. AIR WAY ORAL SET 60/80/100 MM

Specifications

Qualitative Requirements

1. Size: set of 60 mm/70 mm/ 80 mm/90 mm/100 mm
2. Material: good quality plastic; disposable airways
3. Colour coded for different sizes;
4. In sets of above sizes in sterilized packs

Picture of Equipment

