

133. B.P. APPARATUS MERCURY

Specifications

Qualitative Requirements

(Adult)Standard Calibration of 09-30 mm Hg and fitted with locking device to prevent Mercury leak.

ISI Marked . Cuff of synthetic material 5"-6" Tube length 20-22"

Cuff to be fixed by metal clip and right and left side to be indicated.

Picture of equipment

