

### 73. **COLLAR STIFF NECK SHORT**

#### **Specifications**

<b>Qualitative Requirements</b>	<b>Picture of F</b>
<p>CERVICAL COLLAR SHORT (AS PER IS: 11569)</p> <ul style="list-style-type: none"><li>a) Rigid cervical collar-extrication type</li><li>b) Material: soft plastic rigid enough to avoid 70% of neck movement and lined inside with soft cushion.</li><li>c) Padded chain rest for application by sweep technique.</li><li>d) Colour coded and measure holes for easy identification and measurement of correct size.</li><li>e) Velcro for easy fixation.</li><li>d) MRI/CT Scan compatible; radio lucent</li><li>e) Tracheal access opening for advanced airway procedure;</li><li>f) Rear panel opening for cervical spine palpation</li><li>g) Latex free and durable</li><li>h) The size of finished c-collar should not be less than 4.5 cm when measurement as per prescribed procedure.</li></ul>	