

142. STERILIZING DRUM

Specifications

Qualitative Requirements	Picture of Equipment
<ol style="list-style-type: none">1. Stainless steel, seamless2. Size 11" x 9"3. Compatible with latest pressure cooker, type autoclave of 12" x 19".	

